



Model 2000 SHR

GENERAL SPECIFICATIONS

CALIBRATION	TRACEABILITY	17025 Traceable to NIST		
	PURCHASING	Charged Separately / Optional		
OPERATING AMBIENT CONDITIONS	TEMPERATURE	18 °C to 28 °C		
	HUMIDITY	Up to 80 % RH		
STORAGE CONDITIONS	TEMPERATURE	-20 °C to 50 °C		
	HUMIDITY	0 % to 95 % RH (non-condensing)		
ALTITUDE	2000 m			
CONTROLLED RANGE	TEMPERATURE	5 °C to 55 °C		
	HUMIDITY@ 18°C	5 % to 95 % RH	HUMIDITY@ 35°C	5 % to 80 % RH
	HUMIDITY@ 23°C	5 % to 95 % RH	HUMIDITY@ 40°C	5 % to 75 % RH
	HUMIDITY@ 30°C	5 % to 85 % RH	HUMIDITY@ 50°C	5 % to 60 % RH
CALIBRATION SYSTEM ACCURACY	TEMPERATURE	+/- 0.10 °C Or Better	Typically +/- 0.05 °C	
	HUMIDITY	+/- 1.00 % RH Or Better	Specially tuned systems can be as good as ± 0.60 %	
	CONTROLLER TYPE	PID Controller		
UNIFORMITY	TEMPERATURE	0.10 °C Or Better	Typically +/- 0.05 °C	
	HUMIDITY	0.30 % RH @ 18 °C to 28 °C Or Better Typically +/- 0.25 %		
STABILITY	TEMPERATURE	0.05 °C		
	HUMIDITY	0.15 % RH @ 18 °C to 28 °C		
RAMP / SOAK RATE OF CHANGE *Depends on ambient conditions	Temperature	High → Low	1.50 °C / minute (Typical)	
		Low → High	5.00 °C / minute (Typical)	
	Humidity	High → Low	5.00 % RH / minute (Typical)	
		Low → High	10.00 % RH / minute (Typical)	
WARM-UP PERIOD	Ambient Condition	30 Minutes Maximum		
	Cold	30 Minutes		
RESOLUTION	DISHRLAY	Temperature	0.01 °C	
		Humidity	0.01 % RH	
		Dew Point	0.01 °C (Calculated)	
	USB	Temperature	0.01 °C	
		Humidity	0.01 % RH	

CONSUMABLES

WATER	RESERVOIR	200 ml
	SPILL RESISTANT	Yes
	REQUIRED FLUID	Distilled Water Only
	EST REFILL PERIOD	15 Days to 1 Month (Typical) * Depends on Usage
	FILL INDICATOR	Floating Ball
DESICCANT	TYPE	Molecular Sieve
	REPLACEMENT	When Indicating Desiccant is 3/4 Used
	REPLACEMENT FREQUENCY	Depends Entirely on User Workload
	LOCATION	Left Side Mounted
	FASTENER	Desiccant Mounting Bracket Provided
RECALIBRATION	FREQUENCY	Depends on User Uncertainty Requirements
		Once Per Year Recommended

MECHANICAL

GENERATOR DIMENSIONS	Measurement Type	Width	Depth	Height
	Metric	55.33 cm	40.65 cm	24.65 cm
	Imperial	21.78 in	16.02 in	9.70 in
CHAMBER DIMENSIONS	Measurement Type	Diameter	Depth	
	Metric	14.50 cm	20.00 cm	
	Imperial	5.71 in	7.87 in	
WORKING DIMENSIONS AND VOLUME	Measurement Type	Diameter	Depth	
	Metric	12.50 cm	10.90 cm	
	Imperial	4.92 in	4.29 in	
	Volume	1.5 Liter Effective Working Volume		
WEIGHT	Unit Only	Metric	15.5 kg	
		Imperial	34.0 lb	
POWER SUPPLY	Power Supply	12 Volt DC @ 1A		
STANDARD PORT QUANTITY	Depends on doors	Availability: 6 Ports, 5 Ports, 4 Ports, 2 Ports		
CHILLED MIRROR PORTS	In and Out			
PROBE ANALOG OUTPUTS AVAILABLE	0 - 1 Volt	Probe Range		Actual Unit Range
		Temperature	- 40.00 to + 60.00 °C	+ 5.00 to + 60.00 °C
		Humidity	0 to + 100.00 %	+ 5.00 to + 95.00 %

Ordering Information

Standard Supply Package

Square 5-Port Door	Desiccant Canister	Fill Syringe
7 Grommets (1/8",1/4", 3/8", 1/2", 5/8", 3/4", 7/8")	Mains Power Cord	Calibration Documents

Accessories

Doors	Shape		Port Quantity		Connector Type		Part Number	
	Square		5-Port Door		M36 Connectors		01-001-00-0013	
Grommets	Port Type	Size (in)	Size (mm)	Part Number	Port Type	Size (in)	Size (mm)	Part Number
	M36	0.000"	0.000	01-004-09-0001	SLIMLINE	0.125"	3.175	01-004-09-0020
	M36	0.125"	3.175	01-004-09-0002	SLIMLINE	0.250"	6.350	01-004-09-0021
	M36	0.250"	6.350	01-004-09-0003	SLIMLINE	0.375"	9.525	01-004-09-0022
	M36	0.375"	9.525	01-004-09-0004	SLIMLINE	0.500"	12.700	01-004-09-0023
	M36	0.500"	12.700	01-004-09-0005	SLIMLINE	0.625"	15.875	01-004-09-0024
	M36	0.625"	15.875	01-004-09-0006	SLIMLINE	0.750"	19.050	01-004-09-0025
	M36	0.750"	19.050	01-004-09-0007	SLIMLINE	0.825"	20.950	01-004-09-0026
	M36	0.875"	22.225	01-004-09-0008	SLIMLINE	1.000"	25.400	01-004-09-0027
	M36	1.000"	25.400	01-004-09-0009				
Assorted	Description			Part Number		Description		
	Desiccant Canister			01-200-11-0001		Insert		
						01-200-08-0025		



GEO Calibration



Proudly Made in the USA 

2190 Smithtown Ave, Ronkonkoma NY 11779, USA

E: Sales@GeoCalibration.com W: www.GeoCalibration.com T: (631) 471 - 6157

© 2019 GEO Calibration Inc. All rights reserved.