



GEO

Calibration

Model 2015 TS

GENERAL SPECIFICATIONS

CALIBRATION	TRACEABILITY	17025 Traceable to NIST	
	PURCHASING	Charged Separately / Optional	
OPERATING AMBIENT CONDITIONS	TEMPERATURE	18 °C to 28 °C	
	HUMIDITY	Up to 80 % RH	
STORAGE CONDITIONS	TEMPERATURE	-20 °C to 50 °C	
	HUMIDITY	0 % to 95 % RH (non-condensing)	
ALTITUDE	2000 m		
CONTROLLED RANGE	TEMPERATURE	5 °C to 55 °C	
	HUMIDITY@ 18°C	5 % to 95 % RH	HUMIDITY@ 35°C 5 % to 80 % RH
	HUMIDITY@ 23°C	5 % to 95 % RH	HUMIDITY@ 40°C 5 % to 75 % RH
	HUMIDITY@ 30°C	5 % to 85 % RH	HUMIDITY@ 50°C 5 % to 60 % RH
CALIBRATION SYSTEM ACCURACY	TEMPERATURE	+/- 0.10 °C Or Better	Typically +/- 0.05 °C
	HUMIDITY	+/- 1.00 % RH Or Better	Specially tuned systems can be as good as ± 0.60 %
	CONTROLLER TYPE	PID Controller	
UNIFORMITY	TEMPERATURE	0.10 °C Or Better	Typically +/- 0.05 °C
	HUMIDITY	0.30 % RH @ 18 °C to 28 °C Or Better	Typically +/- 0.25 %
STABILITY	TEMPERATURE	0.05 °C	
	HUMIDITY	0.15 % RH @ 18 °C to 28 °C	
RAMP / SOAK RATE OF CHANGE <small>*Depends on ambient conditions</small>	Temperature	High → Low	1.50 °C / minute (Typical)
		Low → High	5.00 °C / minute (Typical)
	Humidity	High → Low	5.00 % RH / minute (Typical)
		Low → High	10.00 % RH / minute (Typical)
WARM-UP PERIOD	Ambient Condition	30 Minutes Maximum	
	Cold	30 Minutes	
RESOLUTION	DISHRLAY	Temperature	0.01 °C
		Humidity	0.01 % RH
		Dew Point	0.01 °C (Calculated)
	USB	Temperature	0.01 °C
		Humidity	0.01 % RH

CONSUMABLES

WATER	RESERVOIR	200 ml
	SPILL RESISTANT	Yes
	REQUIRED FLUID	Distilled Water Only
	EST REFILL PERIOD	15 Days to 1 Month (Typical) * Depends on Usage
	FILL INDICATOR	Floating Ball
DESICCANT	TYPE	Molecular Sieve
	REPLACEMENT	When Indicating Desiccant is 3/4 Used
	REPLACEMENT FREQUENCY	Depends Entirely on User Workload
	LOCATION	Left Side Mounted
	FASTENER	Desiccant Mounting Bracket Provided
RECALIBRATION	FREQUENCY	Depends on User Uncertainty Requirements Once Per Year Recommended

MECHANICAL

	Measurement Type	Width	Depth	Height
GENERATOR DIMENSIONS	Metric	49.65 cm	43.54 cm	23.47 cm
	Imperial	19.55 in	17.14 in	9.24 in
	Measurements Type	Diameter	Depth	
CHAMBER DIMENSIONS	Metric	14.50 cm	20.00 cm	
	Imperial	5.71 in	7.87 in	
	Measurements Type	Diameter	Depth	
WORKING DIMENSIONS AND VOLUME	Metric	12.50 cm	10.90 cm	
	Imperial	4.92 in	4.29 in	
	Volume	1.5 Liter Effective Working Volume		
WEIGHT	Unit Only	Metric	11.34 kg	
		Imperial	25.00 lb	
POWER SUPPLY	Power Supply	12 Volt DC @ 1A		
STANDARD PORT QUANTITY	Depends on doors	Availability: 6 Ports, 5 Ports, 4 Ports, 2 Ports		
CHILLED MIRROR PORTS	In and Out			
		Probe Range	Actual Unit Range	
PROBE ANALOG OUTPUTS AVAILABLE	0 - 1 Volt	Temperature	- 40.00 to + 60.00 °C	+ 5.00 to + 60.00 °C
		Humidity	0 to + 100.00 %	+ 5.00 to + 95.00 %

Ordering Information

Standard Supply Package

Square 6-Port Door	Desiccant Canister	Fill Syringe
7 Grommets (1/8", 1/4", 3/8", 1/2", 5/8", 3/4", 7/8")	Mains Power Cord	Calibration Documents

Accessories

Doors	Shape	Port Quantity	Connector Type	Part Number				
	Square / Round	6-Port Door	M36 Connectors	01-001-00-0017 / 01-001-00-0018				
	Square / Round	4-Port Door	M36 Connectors	01-001-00-0003 / 01-001-00-0004				
	Square / Round	4-Port Door	SLIMLINE	01-001-00-0024 / 01-001-00-0025				
	Square	4-Port Chilled MirrorDoor	M36 Connectors	01-001-00-0010				
	Square	2-Port Door	M36 Connectors	01-001-00-0028				
Grommets	Port Type	Size (in)	Size (mm)	Part Number	Port Type	Size (in)	Size (mm)	Part Number
	M36	0.000"	0.000	01-004-09-0001	SLIMLINE	0.125"	3.175	01-004-09-0020
	M36	0.125"	3.175	01-004-09-0002	SLIMLINE	0.250"	6.350	01-004-09-0021
	M36	0.250"	6.350	01-004-09-0003	SLIMLINE	0.375"	9.525	01-004-09-0022
	M36	0.375"	9.525	01-004-09-0004	SLIMLINE	0.500"	12.700	01-004-09-0023
	M36	0.500"	12.700	01-004-09-0005	SLIMLINE	0.625"	15.875	01-004-09-0024
	M36	0.625"	15.875	01-004-09-0006	SLIMLINE	0.750"	19.050	01-004-09-0025
	M36	0.750"	19.050	01-004-09-0007	SLIMLINE	0.825"	20.950	01-004-09-0026
	M36	0.875"	22.225	01-004-09-0008	SLIMLINE	1.000"	25.400	01-004-09-0027
	M36	1.000"	25.400	01-004-09-0009				
Expanded Vol	Description	Usable Vol:	Width	Depth	Height	Part Number		
	Standard Expansion Chamber		220mm	286mm	345mm	01-054-00-0000		
Assorted	Description	Part Number	Description	Part Number				
	Desiccant Canister	01-200-11-0001	Insert	01-200-08-0025				
	Desiccant Clamp	01-200-01-0013						



GEO Calibration



Proudly Made in the USA



2190 Smithtown Ave, Ronkonkoma NY 11779, USA

E: Sales@GeoCalibration.com W: www.GeoCalibration.com T: (631) 471 - 6157

© 2019 GEO Calibration Inc. All rights reserved.